

Run Report for Instrument: 177

Run By: admin
Method Name: ngmselect_acccredit
Run ID: 201310111024
Reaction Volume: 25 µl
Cover Temperature: 105.0 °C
Total Run Time: 03:29:47
Start Time: 2013-10-11 10:25:25 AM
Finish Time: 2013-10-11 01:55:12 PM
Method Summary: 95.0C 11:00, [94.0C 0:20, 59.0C 3:00]X29, 60.0C 10:00, 4.0C forever
Notes: seconda amplificazione traccia I

VERIFICATION DETAILS

Last TNU Verification:	2013-07-01 01:19:39 PM	PASS
Last Cycle Performance Verification:	2013-07-01 02:36:53 PM	PASS
Last Heated Cover Verification:	2013-07-01 11:59:19 AM	PASS
Last Temperature Verification:	2013-07-01 02:20:09 PM	PASS

RUN ERRORS (0 Errors)

RUN EVENTS (2 Events)

2013-10-11 11:19:14 AM: Run Started
2013-10-11 01:55:12 PM: Run Aborted